

# ***Comparison of Keratometry and Total Keratometry between, IOL Master 700 and Cassini Ambient Topographer***



Ivan Gabric, MD  
Svjetlost Ophthalmology  
*Cassini Ambient & CATALYS cOS6.0 user*

*Presented: During annual ASCRS 2023 in San Diego.*

## **Title:**

Comparison of Keratometry and TOTAL Keratometry between CZM, IOL Master 700 and Cassini Technologies, Cassini Ambient Topographer.

## **Purpose:**

To evaluate level of agreement between these two devices often used in premium lens surgery.

## **Methods:**

Case series of 91 eyes were recorded on both devices by the same operator. Both devices made 3 measurements and a mean of the 3 measurements was used for final analysis. K1, K2, axis of K2, TK1, TK2, axis of TK2 were analyzed.

## **Results:**

The 91 analyzed eyes showed the following data. IOL M group K1: 42.41 SD 1.79, K2: 43.26 SD 1.86, K2 axis 87.4 SD 50.88 TK1 42.43 SD 1.82, K2: 43.32 SD 1.91, K2 axis 92.10 SD 56.34 Cassini group: K1: 42.4 SD 1.78, K2 43.24 SD 1.82 K2 axis 87.1 SD 46.55 TK1 42.54 SD 1.8, K2 43.47, SD 1.84, K2 axis 85.31 SD 51.5 K1 values showed no real difference P value 0.88, K2 values no real difference P value 0.94, TK1 P=0.7, TK 2 P=0.73, Axis of K2 showed a P value of 0.89, TK2 axis also showed a statistically insignificant mismatch of P=0.2

## **Conclusion:**

Both devices offer a very reliable way of obtaining keratometry values for cataract and refractive lens surgery. Keratometry values can be interchanged and used for calculations. Benefits of Cassini over IOL Master 700 is the ability to present topography data beyond the central 4-5 mm area.